





EZZI-LIFT BODY PROTOCOL GUIDE

Thigh and Hip (Thigh 1)

PROTOCOL

| Treatment Type | Treatment Time | What You Will Need |
|----------------|----------------|---|
| Commercial use | 50 minutes | 2 (a)  + 8 (b)  (a) Ezzi-lift microcurrent device (b) Adhesive electrode pads |
| Personal use | 100 minutes | 1 (a)  + 4 (b)  (a) Ezzi-lift microcurrent device (b) Adhesive electrode pads |

*Lead wires are needed to connect pads (sold separately).



1. PLACE THE PADS

Place 4 pads, per device, along the hips and thighs in the pattern shown in the top box. **Pads with the same number attach to a single lead wire.**



2. SELECT A MODE

Select Face Stimulation mode. Adjust power to a comfortable setting. Run Face Stimulation for 25 minutes. After 25 minutes, turn off microcurrent device and remove pads.





3. REPOSITION THE PADS

Reposition 4 pads, per device, along the interior thigh as shown in the bottom box. **Pads with the same number attach to a single lead wire.**



4. SELECT A MODE

Select Face Stimulation mode. Re-adjust power to a comfortable setting. Run Face Stimulation for 25 minutes. After 25 minutes, turn off microcurrent device and remove pads.



REMINDER: New pads must be used for each client. However, a client's personal pads may be used again for that client. Use a marker to put client's name on pads.

MKT-220404-02 A

AVAZIA

